

ALBUMIN (Microalbumin)

Diagnostic reagent for determination of albumin (urine) concentration.

Liquid. Dual Reagents. Store at 2°C - 8°C. For in Vitro Diagnostic Use. Do not freeze

Ref No	Pack	Ref No	Pack	Ref No	Pack	Ref No	Pack
A2290	5 x 25 mL	DMA70	1309 Tests	RMA70	1705 Tests	BY700	2045 Tests
NMA70	647 Tests	MMA70	1675 Tests	RMA71	568 Tests	KMA70	2857 Tests
NMA71	324 Tests	MMA71	837 Tests	LMA70	2000 Tests	TMA70	1842 Tests
SMA70	900 Tests						

INTENDED USE

The test is applied for the quantitative determination of albumin (urine) concentration in urine.

Urinary albumin concentration values provide a good indicator of changes in glomerular permeability, as occur in a number of renal disease.

Diabetic nephropathy is characterized by an early hyperfiltration stage resulting in small increases in urinary albumin excretion. That is why the measurement of urinary albumin is considered a clinically important indicator of deteriorating renal function in diabetic subjects. Urinary albumin excretion is also monitored in hypertensive patients to identify the development of significant nephropathy.

Clinical diagnosis should not be made on the findings of a single test result, but should integrate both clinical and laboratory data.

TEST PRINCIPLE

Albumin in the urine sample causes agglutination of the latex particles coated with anti-human albumin. The agglutination of the particles is proportional to the albumin concentration and can be measured by turbidimetry.

TEST PARAMETERS

Method : Turbidimetric
Wavelength : 540 nm
Temperature : 37°C
Sample : Urine
Linearity : 0.9 mg/L - 130 mg/L

REAGENT COMPOSITION

Reagent 1:

Borate buffer ≤ 0.12 mol/L,

Sodium azide ≤ 1.00 g/L, pH 10.0.

Reagent 2:

Suspension of latex particles coated with anti-human albumin antibodies,

Sodium azide ≤ 1.00 g/L.

S. Albumin Standard: Human albumin.

Albumin concentration is given on the label. Concentration value is traceable to the Standard Reference Material BCR 470 (Institute for Reference Materials and Measurements, IRMM).

Human serum used in the preparation of the standard has been tested and found to be negative for the presence of antibodies anti-HIV and anti-HCV, as well as for HBs antigen. However, the standard should be handled cautiously as potentially infectious.

REAGENT PREPARATION

Working Reagent: Pour the contents of Reagent B vial into Reagent 1 bottle. Mix thoroughly. Reagents are stable for 15 days at 2-8°C.

Smaller Working Reagent volumes can be prepared by mixing: 1 mL of Reagent 2 + 4 mL of Reagent 1. Shake the Latex vial before pipetting.

Albumin Standard (S): Reconstitute with 1.00 mL of distilled water. Standards are stable for 1 month at 2-8°C.

Working reagents are stable at 2-8°C in case of closed vials and avoiding contamination after preparation.

For manual working procedures; if working reagent will be used; shake the Reagent 2 vial gently before pouring its contents into the Reagent 1 bottle. It is advisable to wash the Reagent 2 vial with a small volume of the prepared mixture in order to completely rinse the vial and avoid any losses.

REAGENT STABILITY AND STORAGE

Store at 2-8°C.

Reagents and Standard are stable until the expiry date shown on the label when stored tightly closed and if contaminations are prevented during their use.

Indications of deterioration:

Reagents: absorbance of the blank over 1.200 at 540 nm.

Standard: Presence of moisture.

Once opened vials (reagent 1) are stable minimum 30 days at 2-8°C at optimum conditions. On board stability is strongly related to auto analyzers cooling specification and carry-over values.

SAMPLE

Urine is collected by standard procedures. Urine should be centrifuged before analysis. Albumin in urine is stable for 7 days at 2-8°C, 2 weeks at 4°C (for timed or 24h samples), maximum 6 days at 4°C (for spot samples) and approximately 5 months at -70°C.

TEST PROCEDURE

Sample Start

There have many ready application procedures dedicated to different kind of photometers and ready manual working process can be supplied on request.

There have many ready application procedures dedicated to different kind of biochemistry auto analyzers can be supplied on request.

Substrate Start

There have many ready application procedures dedicated to different kind of biochemistry auto analyzers can be supplied on request.

CALCULATION

The albumin concentration in the sample is calculated using the following general formula:

$$\frac{(A2 - A1) \text{ Sample}}{(A2 - A1) \text{ Standard}} \times C \text{ Standard} \\ = C \text{ Sample (mg/L)}$$

REFERENCE INTERVAL (NORMAL VALUES) *

Urine,
 Adults: Up to 15 mg/L

*It is recommended that each laboratory establish its own reference range.

QUALITY CONTROL AND CALIBRATION

*Calibration Stability: It is strongly depend of application to auto analyzers and auto analyzers specification. Calibration stability is 25 days.

*Each laboratory should establish its own internal Quality Control scheme and procedures for

corrective action if controls do not recover within the acceptable tolerances.

Quality control is recommended every morning. Calibration is not recommended if QC control values are acceptable. Reagent should be calibrated after lot changes.

PERFORMANCE CHARACTERISTICS

Low linearity: 5 mg/L albumin

High Linearity: 130 mg/L albumin. For higher values dilute sample 1/3 with distilled water and repeat measurement. Linearity may considerably vary depending on the instrument used.

Linearity may considerably vary depending on the instrument used.

Precision Studies:

Repeatability (within run) (Intra-assay)

Mean concentration	CV	n
18 mg/L	2.4%	20
57 mg/L	2.4%	20

Reproducibility (run to run) (Inter-assay)

Mean concentration	CV	n
18 mg/L	5.7%	25
57 mg/L	3.6%	25

Sensitivity (LOD): The detection limit of this test is 3 mg/L.

Trueness: Results obtained with this reagent did not show systematic differences when compared with reference reagents. Details of the comparison experiments are available on request.

Prozone effect: falsely low values are obtained when albumin is present in the sample at a concentration higher than 700 mg/L.

Interferences: Bilirubin (20 mg/dL) does not interfere. Hemoglobin (1 g/L) may interfere. Other drugs and substances may interfere.

These performance characteristics have been obtained using an analyzer. Results may vary if a different instrument or a manual procedure is used.

NOTES

1. For in vitro diagnostic use only. Do not pipette by mouth. Avoid contact with skin and mucous membranes.
2. All the calibrators, controls and some reagents must be considered as human & animal sample, so potentially infectious; all the protection actions must be applied to avoid any potential biological risk.
3. Material safety data sheet will be supplied on request.
4. Exercise the normal precautions required for handling laboratory reagents.
5. After measurements are taken, reagent bottles should cap and kept at 2-8°C. Caps of the reagents bottles cannot be used between two different kind of reagent and between R1&R2.
6. Reagents with different lot numbers should not be interchanged or mixed.
7. The reagents contain sodium azide (< 0.1%) as a preservative.
8. The linearity limit depends on the sample to reagent ratio.

PRECAUTIONS AND WASTE DISPOSAL

This product is made to be used in professional laboratories and by professional operators. Perform the test according to the general "Good Laboratory Practice" (GLP) guidelines.

R36/38 : Irritating to eyes and skin.

S20/21 : When using, do not eat, drink or smoke.

S26 : In case of contact with eyes, rinse immediately with plenty of water and seek medical advice.

S28 : After contact with skin wash immediately with plenty of water.

S36/37/39: Wear suitable protective clothing, gloves and eye/face protection.

S45 : In case of accident or if you feel unwell, seek medical advice immediately.

S56 : Dispose of this material and its container at hazardous or special waste collection point.

S57 : Use appropriate container to avoid environmental contamination.

S61 : Avoid release in environment. Refer to special instructions/safety data sheets.

Please consult local regulations for a correct waste disposal.

ABBREVIATIONS

CV% : Coefficient of Variation Percentage

GLP : Good Laboratory Practice

IU : International Unit

mA : milliabsorbance

mL : milliliter





NCCLS: National Committee for Clinical Laboratory Standards

QC : Quality Control

REFERENCES

1. Young DS. Effects of Drugs on Clinical Laboratory Tests. 5th ed. Washington: AACC Press (2000).
2. Friedman and Young. Effects of disease on clinical laboratory tests, 3th ed. AACC Press, 1997.
3. Tietz Textbook of Clinical Chemistry, 3rd edition. Burtis CA, Ashwood ER. WB Saunders Co. 1999.
4. Harmoinen A, Ala-Houhala I, Vuorinen P. Rapid and sensitive immunoassay for albumin determination in urine. Clin Chim Acta 1985; 15;149(2-3):269-74
5. Fielding BA, Price DA, Houlton CA. Enzyme immunoassay for urinary albumin. Clin Chem.1983;29:355-357.
6. Cambiaso CL, Collet-Cassart D, Lievens M. Immunoassay of low concentrations of albumin in urine by latex particle counting. Clin Chem 1988; 34(2):416-418
7. Helsing K. Influenced polymers on the antigen-antibody reaction in a continuous flow system. In: Automated Immunoprecipitin Reactions. Colloquium on AIP. Tarrytown, NY: Technicon Inst. Corp.; 1972:17, 798-9.
8. Wu AHB. Tietz Clinical Guide to Laboratory Tests, 4th edition, Saunders Elsevier, St. Louis, MO: 2006:316. Medcalf EA, Newman DJ, Gorman EG, Price CP. Rapid, robust method for measuring low concentrations of albumin in urine. Clin Chem 1990; 36(3):446-449
9. Clinical and Laboratory Standards Institute (formerly NCCLS). Evaluation of Precision Performance of Quantitative Measurement Methods; Approved Guideline - Second Edition. Wayne, PA: Clinical and Laboratory Standards Institute; 2004. NCCLS Document EP05-A2.
10. Burtis CA, Ashwood ER, Brunis DE. Tietz Textbook of Clinical Chemistry and Molecular Diagnostics, 4th edition. St. Louis, MO.: Elsevier Saunders; 2006:886-888.
11. Bernard A, Lauwerys R. Latex immunoassay of urinary albumin. J Clin Chem Clin Biochem 1983; 21(1):25-30

SYMBOLS

IVD	Only for invitro diagnostic use
LOT	Lot of manufacturing
R1	Reagent 1
R2	Reagent 2
CONC	Concentration
INGRED	Reagent Ingredients
REF	Reference Number (Catalog No)
SN	Serial Number
	Expiration date
	Storage temperature interval
	Read the directions
	Biological risk



 **Archem Diagnostics Industry LTD. ŞTİ.**
Organize Sanayi Bölgesi, Mutsan Sanayi Sitesi
M8 Blok No: 48 Başakşehir / ISTANBUL TURKEY
Tlf: + 90 212 444 08 92
Fax: +90 212 629 98 89
info@archem.com.tr
www.archem.com.tr