

CH CONDITIONER

Liquid. Store at +2/+30°C. For in Vitro Diagnostic Use (IVD). Do not freeze.
 Produced for Siemens Atellica CH930.

Ref No	Package
CON01	2 x 1.5 L

INTENDED USE

CH Conditioner is intended for in vitro diagnostic use in washing and conditioning reaction cuvettes using the Atellica Chemistry Analyser.

ACTIVE COMPOSITION

Tergitol 15-S-9.....5%
 Citric acid.....6,4%
 Potassium sorbate..... 0,2%

PREPARATION

CH Conditioner is ready for use.

STABILITY AND STORAGE

Unopened CH Conditioner packages are stable at +2/+30 °C until the expiration date on the label. The onboard stability of opened CH Conditioner packages is 90 days. Dispose of products whose onboard stability has expired by removing them from the analyzer.

TEST PRINCIPLE






CH Conditioner should be used as specified in the Atellica analyzers user manual

WARNINGS AND PRECAUTIONS

IVD: For in Vitro Diagnostic use only.
 Do not use expired reagents.
 Reagents with two different lot numbers should not be interchanged.
 For professional use.
 Follow Good Laboratory Practice (GLP) guidelines.



H318: Causes serious eye damage.
P280: Wear protective gloves/protective clothing/eye protection/face protection.
P305+P351+P338: IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.
P310: Immediately call a POISON CENTER or doctor/physician.

IVD	In Vitro Diagnostic Medical Device
LOT	Lot number
GTIN	Global Trade Item Number
REF	Reference Number
FOR USE WITH	Identifies products to be used together
PRODUCT OF TURKEY	Product of Turkey
	Manufacturer
	Expiration date
	Temperature limitation
	Humidity Limitation
	Consult instructions for use



Archem Sağlık Sanayi ve Tic. A.Ş.
 Mahmutbey Mah. Halkalı Cad. No:124 Kat:4
 Bağcılar/İstanbul/Turkey
 Tel: + 90 212 444 08 92
 Fax: +90 212 629 98 89
 info@archem.com.tr www.archem.com.tr

