

# CUVETTE CONDITIONER

Liquid. Store at +5/+25°C. For in Vitro Diagnostic Use (IVD). Do not freeze.  
 Produced for Siemens Advia Chemistry Analyser.

Ref No	Package
AD2000	1L

*Changes made in the instructions for use are marked as grey.*

## INTENDED USE

Cuvette Conditioner is intended for in vitro diagnostic use in washing and conditioning reaction cuvettes using the Advia Chemistry Analyser.

## ACTIVE COMPOSITION

Potassium Sorbic Acid..... < %0.2  
 Citric Acid..... < %0.07  
 Surfactant

## PREPARATION

Cuvette Conditioner is ready for use.

## STABILITY AND STORAGE

Unopened Cuvette Conditioner packages are stable at +5/+25 °C until the expiration date on the label. The onboard stability of opened Cuvette Conditioner package is 90 days. Dispose of products whose onboard stability has expired by removing them from the analyzer.

## TEST PRINCIPLE

Cuvette Conditioner should be used as specified in the Advia Chemistry Analyzers user manual

## WARNINGS AND PRECAUTIONS

IVD: For in Vitro Diagnostic use only.  
 Do not use expired reagents.  
 Reagents with two different lot numbers should not be interchanged.  
 For professional use.  
 Follow Good Laboratory Practice (GLP) guidelines.











**H318:** Causes serious eye damage.

**P280:** Wear protective gloves/protective clothing/eye protection/face protection.

**P305+P351+P338:** IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

**P310:** Immediately call a POISON CENTER or doctor/physician.

## SYMBOLS

	In Vitro Diagnostic Medical Device
	Lot number
	Global Trade Item Number
	Reference Number
	Identifies products to be used together
	Product of Turkey
	Manufacturer
	Expiration date
	Temperature limitation
	Consult instructions for use



**Archem Sağlık Sanayi ve Tic. A.Ş.**

Mahmutbey Mah. Halkalı Cad. No:124 Kat:4

Bağcılar/İstanbul/Turkey

Tel: + 90 212 444 08 92

Fax: +90 212 629 98 89

info@archem.com.tr www.archem.com.tr

